
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000715
8:26 21/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:29
PRESSURIZING 3:31
STERILIZING 4:36
136.0 C 229kPa 4:36
136.3 C 232kPa 5:06
136.4 C 233kPa 5:36
136.2 C 230kPa 6:06
136.4 C 232kPa 6:36
136.3 C 231kPa 7:06
136.4 C 232kPa 7:36
136.4 C 232kPa 8:06
136.4 C 232kPa 8:06
VENTING 8:07
AIR DRYING 9:00
CYCLE COMPLETE 9:03

Digital Signature #
73B83B34B59D83F8
