
Midmark M11 - v1.0.5
Total Cycles: 1421

Name: RITTER2
Sterilizer ID: V2313228
08/25/2022 05:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 212.5 F 0.0
2:00 211.6 F 4.6
4:00 221.6 F 11.0
6:00 230.0 F 8.4
8:00 238.4 F 10.3
10:00 243.6 F 12.0
12:00 247.2 F 14.2
14:00 260.4 F 21.6
16:00 272.3 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.4 F 29.5
2:00 271.6 F 28.7
3:00 272.4 F 29.7
4:00 272.1 F 29.1

Min 271.3 F 28.7
Max 272.5 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.0 F 0.0
5:00 224.5 F 0.0
10:00 238.7 F 0.0
15:00 228.8 F 0.0
20:00 223.2 F 0.0
25:00 220.5 F 0.0
30:00 217.7 F 0.0
Drying Complete

FILLING: 2:39
HEATING: 16:30
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:56:14
CYCLE COMPLETE

08/25/2022 06:05 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____