
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007656
7:35 11/07/2024

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:14
PRESSURIZING 3:17
STERILIZING 4:22
136.4 C 220kPa 4:22
136.2 C 222kPa 4:52
135.8 C 219kPa 5:22
136.2 C 221kPa 5:52
136.4 C 219kPa 6:22
136.4 C 222kPa 6:52
136.3 C 222kPa 7:22
136.3 C 220kPa 7:52
136.3 C 220kPa 7:52
VENTING 7:52
AIR DRYING 8:44
CYCLE COMPLETE 38:44

Digital Signature #
DB98B0F6792D7FE5
