
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001046
10:04 29/12/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:57
PRESSURIZING 2:01
STERILIZING 3:02
136.5 C 220kPa 3:02
136.4 C 225kPa 3:32
136.3 C 221kPa 4:02
136.0 C 219kPa 4:32
135.9 C 219kPa 5:02
136.2 C 222kPa 5:32
136.0 C 221kPa 6:02
136.4 C 221kPa 6:32
136.4 C 221kPa 6:32
VENTING 6:32
AIR DRYING 7:25
CYCLE COMPLETE 7:34

Digital Signature #
7876AEF4E2C8BE36
