
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.8uS / 0.5ppm
CYCLE NUMBER 004763
17:08 22/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:08
PRESSURIZING 3:08
STERILIZING 4:07
136.4 C 223kPa 4:07
136.3 C 224kPa 4:37
136.1 C 225kPa 5:07
136.2 C 222kPa 5:37
136.3 C 221kPa 6:07
136.4 C 224kPa 6:37
136.4 C 221kPa 7:07
136.3 C 223kPa 7:37
136.3 C 223kPa 7:37
VENTING 7:37
AIR DRYING 8:32
CYCLE COMPLETE 9:03

Digital Signature #
7EFBD1C8EAC2F29C
