

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4354

Name: RITTER2  
Sterilizer ID: V2313228  
02/13/2025 09:25 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.8 F 0.0  
2:00 70.6 F 0.1  
4:00 77.4 F 0.9  
6:00 107.5 F 2.7  
8:00 135.3 F 5.6  
10:00 163.4 F 9.0  
12:00 185.5 F 12.5  
14:00 200.8 F 4.4  
16:00 212.7 F 8.0  
18:00 225.8 F 12.0  
20:00 227.2 F 7.3  
22:00 239.0 F 12.2  
24:00 237.3 F 9.9  
26:00 249.5 F 15.9  
28:00 260.3 F 22.0  
30:00 270.4 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.4  
1:00 272.9 F 30.0  
2:00 271.6 F 29.4  
3:00 272.9 F 30.1  
4:00 271.8 F 29.3

Min 271.6 F 29.2  
Max 273.0 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.2 F 0.0  
5:00 189.5 F 0.0  
10:00 192.2 F 0.0  
15:00 196.1 F 0.0  
20:00 200.1 F 0.0  
25:00 193.0 F 0.0  
30:00 189.5 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 30:52  
STERILIZING: 4:00  
VENTING: 3:28  
DRYING: 30:00  
TOTAL CYCLE: 01:09:10  
CYCLE COMPLETE

02/13/2025 10:34 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_