
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
1.1uS / 0.7ppm
CYCLE NUMBER 001131
11:06 21/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:22
PRESSURIZING 3:25
STERILIZING 4:35
136.0 C 221kPa 4:35
136.2 C 223kPa 5:05
136.4 C 223kPa 5:35
136.3 C 224kPa 6:05
136.4 C 226kPa 6:35
136.2 C 224kPa 7:05
136.4 C 226kPa 7:35
136.3 C 223kPa 8:05
136.3 C 223kPa 8:05
VENTING 8:06
AIR DRYING 9:00
CYCLE COMPLETE 11:17

Digital Signature #
88DC23C86087B5C6
