

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2411

Name: RITTER 1  
Sterilizer ID: V2178198  
10/06/2022 01:03 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 160.3 F 0.0  
2:00 156.0 F 1.5  
4:00 167.9 F 4.7  
6:00 196.7 F 9.7  
8:00 220.1 F 5.8  
10:00 229.1 F 8.5  
12:00 237.7 F 13.5  
14:00 239.4 F 10.7  
16:00 236.1 F 8.7  
18:00 250.4 F 15.7  
20:00 262.7 F 22.7  
22:00 273.0 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.9  
1:00 272.2 F 29.0  
2:00 273.1 F 29.5  
3:00 271.9 F 29.0  
4:00 272.1 F 29.1

Min 271.7 F 28.7  
Max 273.1 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.6 F 0.0  
5:00 230.9 F 0.0  
10:00 201.7 F 0.0  
15:00 231.2 F 0.0  
20:00 228.6 F 0.0  
25:00 200.7 F 0.0  
30:00 201.4 F 0.0  
Drying Complete

FILLING: 1:12  
HEATING: 22:14  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:00:37  
CYCLE COMPLETE

10/06/2022 02:03 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_