
Midmark M11 - v1.0.5
Total Cycles: 4036

Name: RITTER2
Sterilizer ID: V2313228
10/17/2024 08:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.7 F 0.0
2:00 70.9 F 0.0
4:00 78.2 F 0.0
6:00 113.4 F 0.0
8:00 159.1 F 0.1
10:00 197.2 F 0.4
12:00 205.6 F 2.2
14:00 219.6 F 5.3
16:00 226.9 F 8.4
18:00 233.3 F 12.1
20:00 232.7 F 7.8
22:00 242.8 F 12.7
24:00 239.8 F 10.5
26:00 236.6 F 8.8
28:00 250.2 F 15.4
30:00 261.1 F 21.6
32:00 271.1 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.5 F 29.0
2:00 271.4 F 28.4
3:00 272.5 F 29.1
4:00 271.8 F 28.3

Min 271.3 F 28.3
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.4 F 0.0
5:00 233.7 F 0.0
10:00 224.5 F 0.0
15:00 174.8 F 0.0
20:00 170.7 F 0.0
25:00 170.2 F 0.0
30:00 170.9 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 32:44
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:10:57
CYCLE COMPLETE

10/17/2024 10:01 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____