
Midmark M11 - v1.0.0
Total Cycles: 2833

Name: RITTER 1
Sterilizer ID: V2178198
01/31/2023 12:31 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.7 F 0.0
2:00 173.5 F 4.0
4:00 185.9 F 8.5
6:00 207.7 F 12.1
8:00 221.4 F 7.5
10:00 235.2 F 12.7
12:00 236.4 F 9.8
14:00 237.7 F 8.6
16:00 248.2 F 14.4
18:00 260.4 F 21.1
20:00 271.6 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.2
1:00 271.8 F 28.8
2:00 272.2 F 28.6
3:00 271.2 F 28.2
4:00 272.3 F 29.1

Min 271.1 F 28.1
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 215.0 F 0.0
5:00 215.9 F 0.0
10:00 238.4 F 0.0
15:00 239.7 F 0.0
20:00 229.7 F 0.0
25:00 223.2 F 0.0
30:00 223.0 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 20:30
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:53
CYCLE COMPLETE

01/31/2023 01:29 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____