
Midmark M11 - v1.0.5
Total Cycles: 696

Name: RITTER2
Sterilizer ID: V2313228
01/13/2022 01:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 131.6 F 0.0
2:00 133.4 F 1.9
4:00 160.2 F 5.2
6:00 187.3 F 9.9
8:00 213.3 F 5.4
10:00 222.8 F 9.0
12:00 235.0 F 10.5
14:00 240.8 F 12.0
16:00 239.9 F 10.8
18:00 253.0 F 17.4
20:00 264.3 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.2
1:00 272.4 F 29.8
2:00 271.0 F 29.1
3:00 272.3 F 30.0
4:00 271.6 F 29.3

Min 271.0 F 29.0
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.6 F 0.0
5:00 239.1 F 0.0
10:00 192.8 F 0.0
15:00 222.3 F 0.0
20:00 218.3 F 0.0
25:00 221.2 F 0.0
30:00 219.8 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 21:59
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:00:29
CYCLE COMPLETE

01/13/2022 02:24 PM
ACCEPTED: X
REJECTED:
Operator: U02