
Midmark M11 - v1.0.5
Total Cycles: 2450

Name: RITTER2
Sterilizer ID: V2313228
06/01/2023 09:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.9 F 0.0
2:00 194.7 F 3.4
4:00 199.5 F 8.4
6:00 227.2 F 9.6
8:00 225.9 F 10.2
10:00 230.5 F 7.5
12:00 240.5 F 11.3
14:00 245.7 F 13.7
16:00 258.7 F 20.9
18:00 270.4 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 272.8 F 30.1
2:00 272.2 F 29.5
3:00 271.5 F 29.3
4:00 272.6 F 30.2

Min 271.4 F 29.1
Max 272.9 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.2 F 0.0
5:00 218.9 F 0.0
10:00 233.8 F 0.0
15:00 231.9 F 0.0
20:00 227.1 F 0.0
25:00 223.3 F 0.0
30:00 221.4 F 0.0
Drying Complete

FILLING: 1:48
HEATING: 18:51
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:57:52
CYCLE COMPLETE

06/01/2023 10:30 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____