
Midmark M11 - v1.0.5
Total Cycles: 1124

Name: RITTER2
Sterilizer ID: V2313228
05/26/2022 01:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.9 F 0.0
2:00 184.1 F 3.1
4:00 209.0 F 8.7
6:00 225.9 F 8.5
8:00 234.9 F 10.3
10:00 233.5 F 8.0
12:00 241.4 F 11.4
14:00 246.8 F 13.9
16:00 259.4 F 21.0
18:00 270.9 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.3 F 29.1
2:00 272.0 F 29.6
3:00 271.5 F 28.8
4:00 272.1 F 29.5

Min 271.1 F 28.8
Max 272.5 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.3 F 0.0
5:00 243.1 F 0.0
10:00 225.3 F 0.0
15:00 223.2 F 0.0
20:00 222.2 F 0.0
25:00 206.8 F 0.0
30:00 214.2 F 0.0
Drying Complete

FILLING: 2:51
HEATING: 18:40
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:40
CYCLE COMPLETE

05/26/2022 02:25 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____