

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1899

Name: RITTER 1  
Sterilizer ID: V2178198  
05/03/2022 08:36 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 143.7 F 0.0  
2:00 142.5 F 0.2  
4:00 147.8 F 2.1  
6:00 171.6 F 4.8  
8:00 195.2 F 8.6  
10:00 211.8 F 13.0  
12:00 221.1 F 6.3  
14:00 234.3 F 11.3  
16:00 235.5 F 8.3  
18:00 241.0 F 10.3  
20:00 246.9 F 13.4  
22:00 259.1 F 20.0  
24:00 270.0 F 27.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.0  
1:00 272.5 F 28.7  
2:00 271.4 F 27.9  
3:00 272.3 F 28.8  
4:00 272.3 F 28.3

Min 271.2 F 27.8  
Max 272.6 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.8 F 0.0  
5:00 203.2 F 0.0  
10:00 194.1 F 0.0  
15:00 196.3 F 0.0  
20:00 203.7 F 0.0  
25:00 233.0 F 0.0  
30:00 234.0 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 24:56  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:02:54  
CYCLE COMPLETE

05/03/2022 09:43 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02