
Midmark M11 - v1.0.5
Total Cycles: 472

Name: RITTER2
Sterilizer ID: V2313228
10/22/2021 08:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 103.3 F 0.0
2:00 105.0 F 0.1
4:00 113.4 F 1.2
6:00 139.7 F 4.7
8:00 174.5 F 8.6
10:00 204.0 F 13.0
12:00 224.6 F 7.9
14:00 231.4 F 12.8
16:00 236.2 F 9.8
18:00 235.4 F 9.0
20:00 247.4 F 14.2
22:00 258.7 F 20.5
24:00 269.0 F 27.2
26:00 273.0 F 30.1

STERILIZING
mm:ss Degrees PSI
0:00 272.2 F 30.2
1:00 272.6 F 30.4
2:00 272.4 F 30.8
3:00 271.9 F 30.2
4:00 272.3 F 30.6

Min 271.5 F 30.1
Max 273.3 F 30.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.2 F 0.0
5:00 214.5 F 0.0
10:00 203.0 F 0.0
15:00 203.9 F 0.0
20:00 211.5 F 0.0
25:00 213.0 F 0.0
30:00 219.7 F 0.0

Drying Complete

FILLING: 1:01
HEATING: 26:07
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:04:13
CYCLE COMPLETE

10/22/2021 10:09 AM
ACCEPTED: X
REJECTED:
Operator: U02