

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3092

Name: RITTER 1  
Sterilizer ID: V2178198  
04/11/2023 01:48 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 183.7 F 0.0  
2:00 182.2 F 3.9  
4:00 190.9 F 8.2  
6:00 214.8 F 11.9  
8:00 225.9 F 8.9  
10:00 234.3 F 10.3  
12:00 240.9 F 11.6  
14:00 238.2 F 9.8  
16:00 251.9 F 16.6  
18:00 263.7 F 23.4  
20:00 271.4 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.6  
1:00 272.5 F 29.2  
2:00 271.3 F 28.5  
3:00 272.7 F 29.4  
4:00 271.9 F 28.8  
  
Min 271.2 F 28.4  
Max 272.7 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.8 F 0.0  
5:00 226.1 F 0.0  
10:00 239.6 F 0.0  
15:00 236.5 F 0.0  
20:00 227.7 F 0.0  
25:00 225.3 F 0.0  
30:00 222.4 F 0.0  
Drying Complete

FILLING: 1:26  
HEATING: 20:04  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:58:46  
CYCLE COMPLETE

04/11/2023 02:46 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_