
Midmark M11 - v1.0.0
Total Cycles: 1569

Name: RITTER 1
Sterilizer ID: V2178198
01/26/2022 04:31 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 139.4 F 0.0
2:00 140.5 F 2.2
4:00 157.7 F 5.6
6:00 181.3 F 10.0
8:00 212.0 F 5.5
10:00 218.6 F 8.2
12:00 232.3 F 13.3
14:00 236.7 F 10.0
16:00 236.8 F 8.1
18:00 248.3 F 14.4
20:00 259.8 F 20.8
22:00 270.9 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.6 F 29.1
2:00 272.2 F 28.7
3:00 271.2 F 28.3
4:00 272.3 F 29.1

Min 271.2 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.2 F 0.0
5:00 223.6 F 0.0
10:00 225.3 F 0.0
15:00 227.0 F 0.0
20:00 233.3 F 0.0
25:00 228.9 F 0.0
30:00 225.5 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 22:48
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:01:10
CYCLE COMPLETE

01/26/2022 05:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____