

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000626  
14:29 14/10/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:22  
PRESSURIZING 3:23  
STERILIZING 4:28  
136.5 C 223kPa 4:28  
136.2 C 223kPa 4:58  
136.3 C 223kPa 5:28  
136.4 C 221kPa 5:58  
136.4 C 225kPa 6:28  
136.4 C 225kPa 6:58  
136.4 C 224kPa 7:28  
136.4 C 225kPa 7:58  
136.4 C 225kPa 7:58  
VENTING 7:58  
AIR DRYING 8:51  
CYCLE COMPLETE 9:10

-----  
Digital Signature #  
894385F1AD856954  
-----