
Midmark M11 - v1.0.0
Total Cycles: 3960

Name: RITTER 1
Sterilizer ID: V2178198
10/30/2023 04:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.8 F 0.0
2:00 158.5 F 2.7
4:00 175.7 F 6.2
6:00 201.5 F 10.6
8:00 216.8 F 4.4
10:00 229.3 F 9.4
12:00 233.9 F 9.8
14:00 241.5 F 11.8
16:00 239.9 F 10.3
18:00 252.9 F 16.7
20:00 264.2 F 23.3
22:00 270.9 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 272.4 F 28.8
2:00 272.7 F 29.2
3:00 271.4 F 28.7
4:00 271.9 F 29.0

Min 271.4 F 28.6
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.3 F 0.0
5:00 229.1 F 0.0
10:00 234.5 F 0.0
15:00 227.3 F 0.0
20:00 223.0 F 0.0
25:00 221.6 F 0.0
30:00 204.6 F 0.0
Drying Complete

FILLING: 1:46
HEATING: 22:39
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:02:06
CYCLE COMPLETE

10/30/2023 05:27 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____