
Midmark M11 - v1.0.0
Total Cycles: 3784

Name: RITTER 1
Sterilizer ID: V2178198
09/15/2023 04:35 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 144.9 F 0.0
2:00 145.1 F 0.0
4:00 168.9 F 2.8
6:00 196.3 F 6.6
8:00 217.4 F 11.3
10:00 221.1 F 6.2
12:00 233.8 F 11.2
14:00 233.4 F 8.2
16:00 242.6 F 11.2
18:00 245.8 F 13.1
20:00 258.1 F 19.5
22:00 269.0 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.0 F 28.3
2:00 272.7 F 28.8
3:00 271.5 F 28.4
4:00 271.9 F 28.3

Min 271.3 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.3 F 0.0
5:00 225.4 F 0.0
10:00 230.0 F 0.0
15:00 229.1 F 0.0
20:00 230.2 F 0.0
25:00 227.8 F 0.0
30:00 225.0 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 23:02
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:01:24
CYCLE COMPLETE

09/15/2023 05:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____