
Midmark M11 - v1.0.5
Total Cycles: 3719

Name: RITTER2
Sterilizer ID: V2313228
06/19/2024 02:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.6 F 0.0
2:00 163.0 F 3.0
4:00 186.8 F 7.3
6:00 206.7 F 12.3
8:00 220.5 F 6.7
10:00 234.0 F 12.1
12:00 235.0 F 9.8
14:00 236.6 F 9.0
16:00 248.7 F 15.3
18:00 260.7 F 22.1
20:00 271.4 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.4
1:00 272.3 F 29.5
2:00 272.4 F 29.9
3:00 271.6 F 29.6
4:00 272.3 F 29.6

Min 271.5 F 29.2
Max 272.6 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.4 F 0.0
5:00 217.0 F 0.0
10:00 222.7 F 0.0
15:00 222.0 F 0.0
20:00 216.4 F 0.0
25:00 229.5 F 0.0
30:00 223.7 F 0.0
Drying Complete

FILLING: 1:37
HEATING: 20:35
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 00:59:42
CYCLE COMPLETE

06/19/2024 03:31 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____