

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 005154  
12:44 27/02/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:03  
PRESSURIZING 3:03  
STERILIZING 4:07  
136.4 C 223kPa 4:07  
136.2 C 221kPa 4:37  
136.4 C 225kPa 5:07  
136.4 C 226kPa 5:37  
136.4 C 225kPa 6:07  
136.4 C 225kPa 6:37  
136.3 C 223kPa 7:07  
136.1 C 224kPa 7:37  
136.1 C 224kPa 7:37  
VENTING 7:37  
AIR DRYING 8:28  
CYCLE COMPLETE 15:50

-----  
Digital Signature #  
F4C2A6CFBDBC4876  
-----