
Midmark M11 - v1.0.5
Total Cycles: 3383

Name: RITTER2
Sterilizer ID: V2313228
02/14/2024 09:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.4 F 0.0
2:00 71.9 F 0.0
4:00 83.7 F 0.0
6:00 107.9 F 1.1
8:00 142.5 F 3.9
10:00 165.4 F 7.5
12:00 189.1 F 11.5
14:00 209.3 F 3.7
16:00 218.3 F 7.5
18:00 229.6 F 11.5
20:00 230.6 F 7.7
22:00 242.8 F 13.1
24:00 241.9 F 11.7
26:00 254.3 F 18.0
28:00 265.3 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.4 F 29.4
2:00 271.7 F 29.3
3:00 272.2 F 29.2
4:00 271.8 F 29.5

Min 271.4 F 28.8
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.9 F 0.0
5:00 212.2 F 0.0
10:00 198.5 F 0.0
15:00 193.8 F 0.0
20:00 204.1 F 0.0
25:00 203.0 F 0.0
30:00 206.3 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 29:48
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:07:50
CYCLE COMPLETE

02/14/2024 10:36 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____