
Midmark M11 - v1.0.0
Total Cycles: 1654

Name: RITTER 1
Sterilizer ID: V2178198
02/22/2022 11:42 AM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	104.0 F	0.0
2:00	109.4 F	0.3
4:00	132.4 F	2.0
6:00	165.4 F	5.6
8:00	192.2 F	10.7
10:00	213.5 F	4.4
12:00	232.0 F	10.2
14:00	232.4 F	7.3
16:00	244.1 F	12.1
18:00	245.7 F	12.8
20:00	254.1 F	14.8

C103: HEATUP MODE
STOP PRESSED
CHAMBER TEMP 256F
CHAMBER PRESS 18.4PSI
** ITEMS NOT STERILE **

02/22/2022 12:03 PM