

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2762

Name: RITTER 1  
Sterilizer ID: V2178198  
01/11/2023 08:19 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 78.3 F 0.0  
2:00 78.7 F 0.0  
4:00 85.0 F 0.0  
6:00 111.8 F 0.0  
8:00 143.2 F 0.0  
10:00 173.1 F 2.5  
12:00 196.9 F 5.6  
14:00 202.9 F 9.2  
16:00 214.5 F 12.9  
18:00 222.7 F 6.7  
20:00 233.7 F 11.0  
22:00 232.4 F 7.2  
24:00 245.0 F 13.1  
26:00 242.4 F 11.3  
28:00 254.5 F 17.5  
30:00 265.1 F 23.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.3  
1:00 272.5 F 28.9  
2:00 271.6 F 28.2  
3:00 272.7 F 29.0  
4:00 271.7 F 28.2  
  
Min 271.5 F 28.1  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.0 F 0.0  
5:00 202.8 F 0.0  
10:00 216.8 F 0.0  
15:00 197.8 F 0.0  
20:00 204.2 F 0.0  
25:00 212.1 F 0.0  
30:00 216.6 F 0.0  
Drying Complete

FILLING: 0:47  
HEATING: 31:55  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:09:52  
CYCLE COMPLETE

01/11/2023 09:28 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_