

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 002522  
10:02 20/02/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:03  
PRESSURIZING 2:10  
STERILIZING 3:18  
136.0 C 221kPa 3:18  
136.4 C 225kPa 3:48  
136.0 C 222kPa 4:18  
136.4 C 225kPa 4:48  
136.4 C 225kPa 5:18  
136.3 C 224kPa 5:48  
136.1 C 223kPa 6:18  
136.2 C 223kPa 6:48  
136.2 C 223kPa 6:48  
VENTING 6:48  
AIR DRYING 7:39  
CYCLE COMPLETE 9:30  
-----

Digital Signature #  
849E90112BB3BF52  
-----