
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000191
15:17 06/07/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:32
PRESSURIZING 4:32
STERILIZING 5:57
136.4 C 222kPa 5:57
136.4 C 226kPa 6:27
136.1 C 219kPa 6:57
136.3 C 223kPa 7:27
136.4 C 223kPa 7:57
136.3 C 225kPa 8:27
135.9 C 219kPa 8:57
136.3 C 224kPa 9:27
136.3 C 224kPa 9:27
VENTING 9:27
AIR DRYING 10:27
CYCLE COMPLETE 10:30

Digital Signature #
B4772DF7B58E8EBE
