
Midmark M11 - v1.0.5
Total Cycles: 3551

Name: RITTER2
Sterilizer ID: V2313228
04/24/2024 08:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.8 F 0.0
2:00 83.8 F 0.1
4:00 134.7 F 0.5
6:00 106.8 F 2.5
8:00 136.8 F 5.2
10:00 164.0 F 8.3
12:00 186.3 F 11.8
14:00 217.1 F 4.2
16:00 216.6 F 7.9
18:00 231.3 F 11.9
20:00 231.2 F 7.7
22:00 240.7 F 12.8
24:00 239.7 F 10.8
26:00 251.8 F 16.7
28:00 262.3 F 22.8
30:00 272.4 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.5
1:00 272.6 F 29.5
2:00 272.5 F 29.8
3:00 272.7 F 30.1
4:00 271.5 F 29.7

Min 271.4 F 29.3
Max 273.0 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.7 F 0.0
5:00 233.3 F 0.0
10:00 182.1 F 0.0
15:00 176.7 F 0.0
20:00 174.3 F 0.0
25:00 172.8 F 0.0
30:00 171.8 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 31:23
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:09:31
CYCLE COMPLETE

04/24/2024 09:32 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____