
Mississauga, ON
Canada

STATIM 5000 S501R751
SN 511020J00004
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000817
11:38 28/03/2022
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:30
PRESSURIZING 7:47
STERILIZING 9:42
136.3 C 221kPa 9:42
136.3 C 222kPa 10:12
136.3 C 222kPa 10:42
136.4 C 223kPa 11:12
136.4 C 220kPa 11:42
136.4 C 223kPa 12:12
136.4 C 224kPa 12:42
136.0 C 221kPa 13:12
136.0 C 221kPa 13:12
VENTING 13:12
AIR DRYING 14:25
CYCLE COMPLETE 44:25

Digital Signature #
D61F9562C7C66950
