
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001547
14:44 24/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:05
PRESSURIZING 3:07
STERILIZING 4:01
136.0 C 220kPa 4:01
136.4 C 225kPa 4:31
136.2 C 224kPa 5:01
136.4 C 225kPa 5:31
136.2 C 224kPa 6:01
136.3 C 225kPa 6:31
136.3 C 225kPa 7:01
136.3 C 224kPa 7:31
136.3 C 224kPa 7:31
VENTING 7:32
AIR DRYING 8:23
CYCLE COMPLETE 12:26

Digital Signature #
E438E9055F8E95C7
