
Midmark M11 - v1.0.0
Total Cycles: 2662

Name: RITTER 1
Sterilizer ID: V2178198
12/08/2022 02:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.5 F 0.0
2:00 189.1 F 3.8
4:00 210.8 F 8.7
6:00 219.9 F 7.7
8:00 232.6 F 9.5
10:00 236.9 F 8.8
12:00 244.3 F 13.0
14:00 243.6 F 11.9
16:00 256.9 F 18.9
18:00 268.6 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.6 F 28.9
2:00 271.7 F 28.1
3:00 272.0 F 28.7
4:00 272.4 F 28.7

Min 271.3 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.7 F 0.0
5:00 239.0 F 0.0
10:00 229.9 F 0.0
15:00 224.3 F 0.0
20:00 223.0 F 0.0
25:00 220.7 F 0.0
30:00 219.6 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 19:11
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:57:59
CYCLE COMPLETE

12/08/2022 03:22 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____