
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005934
10:39 10/07/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:03
PRESSURIZING 2:07
STERILIZING 3:10
136.5 C 220kPa 3:10
135.9 C 222kPa 3:40
136.3 C 224kPa 4:10
136.4 C 225kPa 4:40
136.4 C 222kPa 5:10
136.3 C 223kPa 5:40
136.3 C 224kPa 6:10
136.2 C 222kPa 6:40
136.2 C 222kPa 6:40
VENTING 6:40
AIR DRYING 7:32
CYCLE COMPLETE 9:27

Digital Signature #
1F7F6F198735C8DC
