

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 001925  
7:37 08/08/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:28  
PRESSURIZING 2:36  
STERILIZING 3:47  
136.1 C 220kPa 3:47  
136.4 C 224kPa 4:17  
136.4 C 224kPa 4:47  
136.4 C 225kPa 5:17  
136.2 C 223kPa 5:47  
136.3 C 223kPa 6:17  
136.4 C 224kPa 6:47  
136.1 C 222kPa 7:17  
136.1 C 222kPa 7:17  
VENTING 7:18  
AIR DRYING 8:10  
CYCLE COMPLETE 38:10  
-----

Digital Signature #  
441B0489AB65F5EB  
-----