

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4043

Name: RITTER2  
Sterilizer ID: V2313228  
10/18/2024 02:02 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 87.6 F 0.0  
2:00 89.7 F 0.0  
4:00 116.0 F 0.0  
6:00 150.2 F 0.2  
8:00 172.4 F 1.4  
10:00 187.8 F 4.0  
12:00 198.0 F 6.9  
14:00 209.6 F 10.3  
16:00 219.6 F 9.9  
18:00 226.6 F 7.8  
20:00 236.7 F 12.0  
22:00 235.5 F 9.2  
24:00 241.5 F 11.0  
26:00 246.0 F 13.4  
28:00 256.7 F 19.1  
30:00 266.4 F 25.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.6  
1:00 272.7 F 29.3  
2:00 272.0 F 28.7  
3:00 271.7 F 29.0  
4:00 272.7 F 29.3

Min 271.4 F 28.5  
Max 272.7 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.6 F 0.0  
5:00 190.5 F 0.0  
10:00 188.0 F 0.0  
15:00 183.4 F 0.0  
20:00 180.3 F 0.0  
25:00 180.2 F 0.0  
30:00 183.4 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 31:48  
STERILIZING: 4:00  
VENTING: 3:49  
DRYING: 30:00  
TOTAL CYCLE: 01:10:30  
CYCLE COMPLETE

10/18/2024 03:12 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_