
Midmark M11 - v1.0.0
Total Cycles: 1876

Name: RITTER 1
Sterilizer ID: V2178198
04/27/2022 07:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.7 F 0.0
2:00 79.6 F 0.0
4:00 84.1 F 0.0
6:00 120.4 F 0.0
8:00 152.9 F 1.5
10:00 174.1 F 3.7
12:00 199.2 F 6.4
14:00 205.1 F 10.1
16:00 223.0 F 9.3
18:00 225.0 F 7.2
20:00 235.7 F 11.4
22:00 232.8 F 7.6
24:00 245.4 F 13.4
26:00 243.1 F 11.6
28:00 254.9 F 17.8
30:00 265.4 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.4 F 28.7
2:00 271.5 F 28.3
3:00 272.4 F 28.6
4:00 271.5 F 28.3

Min 271.3 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.9 F 0.0
5:00 227.6 F 0.0
10:00 210.8 F 0.0
15:00 224.0 F 0.0
20:00 231.0 F 0.0
25:00 228.2 F 0.0
30:00 227.8 F 0.0
Drying Complete

FILLING: 0:43
HEATING: 31:53
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:09:45
CYCLE COMPLETE

04/27/2022 08:25 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____