
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.6uS / 0.3ppm
CYCLE NUMBER 000507
16:21 07/05/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:35
PRESSURIZING 2:43
STERILIZING 3:58
136.0 C 229kPa 3:58
136.1 C 231kPa 4:28
136.4 C 233kPa 4:58
136.4 C 233kPa 5:28
136.4 C 233kPa 5:58
136.0 C 231kPa 6:28
136.4 C 233kPa 6:58
136.4 C 233kPa 7:28
136.4 C 233kPa 7:28
VENTING 7:28
AIR DRYING 8:24
CYCLE COMPLETE 13:40

Digital Signature #
3E1685284546C453
