
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
1.1uS / 0.7ppm
CYCLE NUMBER 001533
9:26 10/27/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:22
PRESSURIZING 6:25
STERILIZING 8:05
136.0 C 220kPa 8:05
136.0 C 217kPa 8:35
136.4 C 220kPa 9:05
136.2 C 218kPa 9:35
136.0 C 218kPa 10:05
136.3 C 218kPa 10:35
136.4 C 222kPa 11:06
136.4 C 222kPa 11:35
VENTING 11:35
AIR DRYING 12:39
CYCLE COMPLETE 31:49

Digital Signature #
CF10F0CCC74CECC4
