
Midmark M11 - v1.0.0
Total Cycles: 4709

Name: RITTER 1
Sterilizer ID: V2178198
07/10/2024 10:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.9 F 0.0
2:00 158.2 F 2.4
4:00 162.8 F 5.5
6:00 181.3 F 9.2
8:00 205.6 F 13.4
10:00 218.4 F 5.7
12:00 231.1 F 9.7
14:00 236.5 F 10.5
16:00 238.3 F 10.5
18:00 237.4 F 8.4
20:00 247.9 F 14.3
22:00 258.9 F 20.3
24:00 269.1 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.6 F 29.1
2:00 272.0 F 28.5
3:00 271.7 F 28.6
4:00 272.6 F 29.1

Min 271.5 F 28.2
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.3 F 0.0
5:00 225.9 F 0.0
10:00 206.8 F 0.0
15:00 240.7 F 0.0
20:00 229.9 F 0.0
25:00 229.7 F 0.0
30:00 227.4 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 25:11
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:03:53
CYCLE COMPLETE

07/10/2024 11:59 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____