
Midmark M11 - v1.0.0
Total Cycles: 4054

Name: RITTER 1
Sterilizer ID: V2178198
11/23/2023 11:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 110.8 F 0.0
2:00 110.2 F 0.6
4:00 114.1 F 2.5
6:00 145.6 F 4.9
8:00 189.9 F 8.6
10:00 210.2 F 12.6
12:00 216.0 F 4.3
14:00 224.1 F 8.1
16:00 234.7 F 12.3
18:00 233.3 F 8.7
20:00 242.5 F 12.1
22:00 243.1 F 12.1
24:00 254.5 F 18.0
26:00 265.0 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.7
1:00 272.4 F 29.3
2:00 271.3 F 28.6
3:00 272.7 F 29.7
4:00 272.0 F 29.1

Min 271.0 F 28.6
Max 273.3 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.2 F 0.0
5:00 232.5 F 0.0
10:00 238.2 F 0.0
15:00 227.1 F 0.0
20:00 223.6 F 0.0
25:00 226.3 F 0.0
30:00 225.0 F 0.0

Drying Complete

FILLING: 1:04
HEATING: 27:59
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:06:48
CYCLE COMPLETE

11/23/2023 12:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____