

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.9uS / 0.5ppm  
CYCLE NUMBER 001480  
15:01 10/06/2021  
WRAPPED  
134 C for 6 min.  
CYCLE START 0:00  
CONDITIONING 3:42  
PRESSURIZING 11:46  
STERILIZING 12:53  
136.0 C 220kPa 12:53  
136.3 C 220kPa 13:23  
136.1 C 221kPa 13:53  
136.3 C 222kPa 14:23  
136.5 C 219kPa 14:53  
136.5 C 223kPa 15:23  
136.4 C 222kPa 15:53  
136.2 C 222kPa 16:24  
136.3 C 223kPa 16:54  
136.4 C 220kPa 17:24  
136.4 C 223kPa 17:54  
136.5 C 223kPa 18:24  
136.4 C 223kPa 18:53  
VENTING 18:54  
AIR DRYING 19:52  
CYCLE COMPLETE 49:52  
-----

Digital Signature #  
C7A7C827483B9D2E  
-----