
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000885
17:35 26/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:39
PRESSURIZING 2:50
STERILIZING 3:55
136.0 C 227kPa 3:55
136.1 C 231kPa 4:25
136.3 C 232kPa 4:55
136.4 C 233kPa 5:25
136.3 C 233kPa 5:55
136.2 C 231kPa 6:25
136.2 C 231kPa 6:55
136.4 C 233kPa 7:25
136.4 C 233kPa 7:25
VENTING 7:25
AIR DRYING 8:18
CYCLE COMPLETE 23:18

Digital Signature #
0D684CBAF2569513
