
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000936
10:36 17/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:38
PRESSURIZING 2:40
STERILIZING 3:50
136.0 C 228kPa 3:50
136.1 C 230kPa 4:20
136.1 C 231kPa 4:50
136.3 C 231kPa 5:20
136.4 C 233kPa 5:50
136.4 C 231kPa 6:20
136.3 C 232kPa 6:50
136.4 C 233kPa 7:20
136.4 C 233kPa 7:20
VENTING 7:20
AIR DRYING 8:14
CYCLE COMPLETE 23:14

Digital Signature #
D0F26C363EAD2311
