
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004474
16:48 03/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:00
PRESSURIZING 3:03
STERILIZING 4:09
136.4 C 220kPa 4:09
136.3 C 224kPa 4:39
136.4 C 222kPa 5:09
136.4 C 224kPa 5:39
136.3 C 219kPa 6:09
136.4 C 225kPa 6:39
136.5 C 225kPa 7:09
136.4 C 223kPa 7:39
136.4 C 223kPa 7:39
VENTING 7:39
AIR DRYING 8:42
CYCLE COMPLETE 8:44

Digital Signature #
3F84EC31FCFDFFF4
