

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4066

Name: RITTER2  
Sterilizer ID: V2313228  
10/30/2024 08:50 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.9 F 0.0  
2:00 71.0 F 0.0  
4:00 78.4 F 0.0  
6:00 118.2 F 0.1  
8:00 151.0 F 0.2  
10:00 178.9 F 0.4  
12:00 196.9 F 3.0  
14:00 210.6 F 6.3  
16:00 221.0 F 10.0  
18:00 229.6 F 9.8  
20:00 234.8 F 9.7  
22:00 237.8 F 10.0  
24:00 244.3 F 12.9  
26:00 243.0 F 11.9  
28:00 255.2 F 18.2  
30:00 266.0 F 24.8

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 28.6  
1:00 272.7 F 29.3  
2:00 271.5 F 28.6  
3:00 272.4 F 29.0  
4:00 271.7 F 28.9

Min 271.5 F 28.4  
Max 272.7 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.1 F 0.0  
5:00 211.4 F 0.0  
10:00 206.1 F 0.0  
15:00 195.1 F 0.0  
20:00 190.0 F 0.0  
25:00 189.6 F 0.0  
30:00 199.6 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 31:33  
STERILIZING: 4:00  
VENTING: 3:27  
DRYING: 30:00  
TOTAL CYCLE: 01:09:49  
CYCLE COMPLETE

10/30/2024 09:59 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_