
Midmark M11 - v1.0.0
Total Cycles: 3905

Name: RITTER 1
Sterilizer ID: V2178198
10/16/2023 10:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.3 F 0.0
2:00 169.5 F 1.6
4:00 170.3 F 4.5
6:00 183.6 F 8.6
8:00 204.9 F 13.3
10:00 225.6 F 6.9
12:00 233.3 F 11.4
14:00 233.3 F 7.8
16:00 245.3 F 13.2
18:00 242.7 F 11.6
20:00 254.2 F 17.5
22:00 264.4 F 23.8
24:00 271.0 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.3 F 29.0
2:00 272.1 F 29.2
3:00 271.9 F 28.7
4:00 272.6 F 29.3

Min 271.4 F 28.5
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.7 F 0.0
5:00 237.9 F 0.0
10:00 220.5 F 0.0
15:00 221.5 F 0.0
20:00 233.3 F 0.0
25:00 226.4 F 0.0
30:00 231.0 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 24:40
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:03:39
CYCLE COMPLETE

10/16/2023 11:05 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____