

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5296

Name: RITTER 1  
Sterilizer ID: V2178198  
01/29/2025 06:58 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 185.5 F 0.0  
2:00 178.9 F 4.5  
4:00 200.5 F 9.8  
6:00 216.8 F 5.5  
8:00 235.0 F 10.3  
10:00 231.5 F 6.9  
12:00 244.3 F 12.3  
14:00 244.4 F 12.4  
16:00 258.0 F 19.6  
18:00 269.9 F 27.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.3  
1:00 272.4 F 28.5  
2:00 272.6 F 29.0  
3:00 271.7 F 28.1  
4:00 272.5 F 28.9  
  
Min 271.5 F 28.1  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.3 F 0.0  
5:00 224.4 F 0.0  
10:00 221.7 F 0.0  
15:00 231.5 F 0.0  
20:00 241.3 F 0.0  
25:00 234.0 F 0.0  
30:00 227.3 F 0.0  
  
Drying Complete

FILLING: 1:53  
HEATING: 18:51  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:58:00  
CYCLE COMPLETE

01/29/2025 07:55 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_