
Midmark M11 - v1.0.0
Total Cycles: 3965

Name: RITTER 1
Sterilizer ID: V2178198
11/01/2023 10:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.7 F 0.0
2:00 150.5 F 2.7
4:00 173.5 F 6.2
6:00 192.8 F 10.2
8:00 218.7 F 5.3
10:00 224.7 F 7.8
12:00 239.0 F 12.5
14:00 235.6 F 9.0
16:00 240.2 F 10.4
18:00 245.4 F 12.8
20:00 257.3 F 19.1
22:00 268.0 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.5
1:00 272.6 F 29.2
2:00 272.0 F 28.6
3:00 272.3 F 29.1
4:00 272.9 F 28.9

Min 271.5 F 28.4
Max 273.1 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.5 F 0.0
5:00 210.2 F 0.0
10:00 232.5 F 0.0
15:00 228.8 F 0.0
20:00 218.7 F 0.0
25:00 213.9 F 0.0
30:00 201.0 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 23:58
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:02:35
CYCLE COMPLETE

11/01/2023 11:37 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____