

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2775

Name: RITTER 1  
Sterilizer ID: V2178198  
01/12/2023 01:08 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 207.9 F 0.0  
2:00 206.8 F 0.9  
4:00 214.2 F 6.5  
6:00 228.7 F 12.1  
8:00 231.2 F 8.5  
10:00 239.3 F 10.6  
12:00 244.4 F 13.0  
14:00 244.0 F 12.3  
16:00 257.7 F 19.6  
18:00 269.8 F 27.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.4  
1:00 272.6 F 29.0  
2:00 271.7 F 28.3  
3:00 272.5 F 29.2  
4:00 272.4 F 28.8  
  
Min 271.5 F 28.2  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 214.7 F 0.0  
5:00 223.9 F 0.0  
10:00 233.7 F 0.0  
15:00 228.8 F 0.0  
20:00 229.0 F 0.0  
25:00 230.3 F 0.0  
30:00 227.8 F 0.0  
  
Drying Complete

FILLING: 1:39  
HEATING: 18:56  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:57:51  
CYCLE COMPLETE

01/12/2023 02:05 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_