
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001385
11:17 16/12/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:37
PRESSURIZING 3:39
STERILIZING 4:48
136.0 C 220kPa 4:48
136.4 C 225kPa 5:18
135.9 C 220kPa 5:48
136.3 C 224kPa 6:18
136.4 C 225kPa 6:48
136.4 C 224kPa 7:18
136.4 C 225kPa 7:48
136.4 C 225kPa 8:18
136.4 C 225kPa 8:18
VENTING 8:18
AIR DRYING 9:13
CYCLE COMPLETE 15:23

Digital Signature #
F905D68B97BBC871
