
Midmark M11 - v1.0.5
Total Cycles: 1632

Name: RITTER2
Sterilizer ID: V2313228
10/20/2022 12:33 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.0 F 0.0
2:00 185.2 F 3.6
4:00 194.4 F 8.6
6:00 215.8 F 8.9
8:00 228.8 F 9.8
10:00 233.0 F 8.7
12:00 242.1 F 12.1
14:00 245.1 F 13.5
16:00 258.2 F 20.6
18:00 269.7 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 272.2 F 29.2
1:00 272.4 F 30.0
2:00 272.1 F 29.3
3:00 271.9 F 29.6
4:00 272.5 F 29.8

Min 271.6 F 29.0
Max 272.6 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.4 F 0.0
5:00 205.1 F 0.0
10:00 208.9 F 0.0
15:00 207.9 F 0.0
20:00 205.6 F 0.0
25:00 203.1 F 0.0
30:00 201.3 F 0.0
Drying Complete

FILLING: 2:53
HEATING: 18:50
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:59:01
CYCLE COMPLETE

10/20/2022 01:31 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____